

VERNIER CALIPER WITH TITANIUM PLATED BEAM



titanium plated beam

1219-150

Graduation 0.02mm

Code	Range	Accuracy
1219-1501	0-150mm	±0.03mm

Graduation 0.02mm/0.001"


Code	Range	Accuracy
1219-1502	0-150mm/0-6"	±0.03mm

Graduation 0.05mm

Code	Range	Accuracy
1219-1503	0-150mm	±0.05mm

Graduation 0.05mm/1/128"

Code	Range	Accuracy
1219-150	0-150mm/0-6"	±0.05mm



- Meet DIN862
- Titanium plated beam for wear resistant and easy moving
- Made of stainless steel
- Satin chrome plated reading surface